

You have not supplied any structure factors. As a result the full set of tests cannot be run.

No syntax errors found. CIF dictionary Interpreting this report

Bond precision:	C-C = 0.0037 A	Wavelength=0.71073		
Cell:	a=11.6451(16)	b=7.2951(8)	c=16.761(2)	
	alpha=90	beta=90	gamma=90	
Temperature:	253 K			

```
Correction method= # Reported T Limits: Tmin=0.657 Tmax=0.746
AbsCorr = MULTI-SCAN
```

```
R(reflections)= 0.0349( 1500)      wR2(reflections)=
S = 1.081                          0.0978( 1631)
Npar= 115
```

---

The following ALERTS were generated. Each ALERT has the format

**test-name\_ALERT\_alert-type\_alert-level.**

Click on the hyperlinks for more details of the test.

---



### Alert level C

PLAT430\_ALERT\_2\_C Short Inter D...A Contact O1 ..01 . 2.87 Ang.  
-x,y,3/2-z = 3\_556 Check

---



### Alert level G

PLAT002\_ALERT\_2\_G Number of Distance or Angle Restraints on AtSite 1 Note  
PLAT004\_ALERT\_5\_G Polymeric Structure Found with Maximum Dimension 2 Info  
PLAT007\_ALERT\_5\_G Number of Unrefined Donor-H Atoms ..... 3 Report  
H1A H1B H3  
PLAT012\_ALERT\_1\_G N.O.K. \_shelx\_res\_checksum Found in CIF ..... Please Check  
PLAT172\_ALERT\_4\_G The CIF-Embedded .res File Contains DFIX Records 1 Report  
PLAT794\_ALERT\_5\_G Tentative Bond Valency for Zn1 (II) . 1.87 Info  
PLAT860\_ALERT\_3\_G Number of Least-Squares Restraints ..... 1 Note  
PLAT883\_ALERT\_1\_G No Info/Value for \_atom\_sites\_solution\_primary . Please Do !

---

- 0 **ALERT level A** = Most likely a serious problem - resolve or explain  
0 **ALERT level B** = A potentially serious problem, consider carefully  
1 **ALERT level C** = Check. Ensure it is not caused by an omission or oversight  
8 **ALERT level G** = General information/check it is not something unexpected
- 2 ALERT type 1 CIF construction/syntax error, inconsistent or missing data  
2 ALERT type 2 Indicator that the structure model may be wrong or deficient  
1 ALERT type 3 Indicator that the structure quality may be low  
1 ALERT type 4 Improvement, methodology, query or suggestion  
3 ALERT type 5 Informative message, check
- 
-

It is advisable to attempt to resolve as many as possible of the alerts in all categories. Often the minor alerts point to easily fixed oversights, errors and omissions in your CIF or refinement strategy, so attention to these fine details can be worthwhile. In order to resolve some of the more serious problems it may be necessary to carry out additional measurements or structure refinements. However, the purpose of your study may justify the reported deviations and the more serious of these should normally be commented upon in the discussion or experimental section of a paper or in the "special\_details" fields of the CIF. checkCIF was carefully designed to identify outliers and unusual parameters, but every test has its limitations and alerts that are not important in a particular case may appear. Conversely, the absence of alerts does not guarantee there are no aspects of the results needing attention. It is up to the individual to critically assess their own results and, if necessary, seek expert advice.

### **Publication of your CIF in IUCr journals**

A basic structural check has been run on your CIF. These basic checks will be run on all CIFs submitted for publication in IUCr journals (*Acta Crystallographica*, *Journal of Applied Crystallography*, *Journal of Synchrotron Radiation*); however, if you intend to submit to *Acta Crystallographica Section C* or *E* or *IUCrData*, you should make sure that full publication checks are run on the final version of your CIF prior to submission.

### **Publication of your CIF in other journals**

Please refer to the *Notes for Authors* of the relevant journal for any special instructions relating to CIF submission.

Datablock 1 - ellipsoid plot

