

Supplementary Table S2: Overview ddPCR assay development ar

Gene	AA change	LOD (MUT:WT)	LOD (%)
<i>EGFR</i>	T790M	1:1122	0.089
<i>EGFR</i>	del19	1:753	0.130
<i>EGFR</i>	L858R	1:1316	0.076
<i>EGFR</i>	C797S	1:2140	0.047
<i>EGFR</i>	L861Q	1:2584	0.039
<i>BRAF</i>	V600E	1:2145	0.047
<i>BRAF</i>	V600K	1:6530	0.015
<i>BRAF</i>	V600R	1:2019	0.050
<i>KRAS</i> *	G12V	1:9041	0.011
<i>NRAS</i>	Q61R	1:994	0.010
<i>KIT</i>	V559A	1:1396	0.072
<i>KIT</i>	D816V	1:2174	0.046
* This assay was developed by Hussung et al., J Mol Diagn, 2020, PM			

nd validation.		
LOB (MUT copies/mL plasma)	Specificity	
3.2	100%	
17.8	100%	
2.7	100%	
1.7	100%	
1.7	100%	
5.9	100%	
3.1	100%	
3.3	100%	
2.2	100%	
6.2	100%	
2.7	100%	
0	100%	
IID: 32376474		