

Supplementary Table S2. Overview and analysis of samples used for ddPCR. The table displays sample numbers and corresponding information; the quantity of extracted cfDNA measured with Qubit and the respective cfDNA amount in 7 µl used per ddPCR assay; the fractional abundance (FA); the measured concentration in copies/µl and the calculated quantity of cfDNA in ng.

sample no.	Sample content	Storage conditions	DNA [ng] with Qubit (25 µl eluate)	DNA [ng] used per ddPCR assay (7µl)	no. of positive mutant droplets (7µl)	FA [%]	DNA conc. [copies/µl] measured by ddPCR	Calculated* * DNA [ng] per ddPCR assay [7 µl]
1	NC	No buffer, 0h	6.02	1.69	-	26.60	1.93	
2	Spike AF 5%	No buffer, 0h	4.16	1.16	1; 0	0	30.2	2.20
3	NC	UAS, 0h	8.89	2.49	-	47.8	3.47	
4	Spike AF 5%	UAS, 0h	17.13	4.80	25; 23	2.05	85.6	6.50
5	NC	UAS, 4h RT	9.99	2.80	NA*			
6	Spike AF 5%	UAS, 4h, RT	19.33	5.41	34; 21	2.00	112.7	8.18
7	NC	UAS, 4h, F	6.99	1.96	-	48.40	3.52	
8	Spike AF 5%	UAS, 4h, F	17.90	5.01	14; 14	3.20	75.95	5.51
9	NC	UAS, 24h, RT	15.40	4.31	-	88.50	6.43	
10	Spike AF 5%	UAS, 24h, RT	25.33	7.09	20	1.90	126.40	9.18
11	NC	UAS, 24h, F	10.40	2.91	-	58.50	4.25	
12	Spike AF 5%	UAS, 24h, F	13.87	3.88	19; 30	1.80	105.20	7.64
13	NC	AlloU, 0h	2.14	0.60	-	8.05	0.58	
14	Spike AF 5%	AlloU, 0h	14.47	4.05	23; 36	2.95	73.1	5.30
15	NC	AlloU, 4h, RT	1.69	0.47	-	6.20	0.45	
16	Spike AF 5%	AlloU, 4h, RT	8.32	2.33	8; 3	2.25	41.70	3.03
17	NC	AlloU, 4h, F	1.87	0.52	-	7.1	0.52	
18	Spike AF 5%	AlloU, 4h, F	11.97	3.35	8; 2	2.5	16.55	1.20
19	NC	AlloU, 24h, RT	3.81	1.07	-	5.35	0.39	
20	Spike AF 5%	AlloU, 24h, RT	8.15	2.28	0	0	32.1	2.33
21	NC	AlloU, 24h, F	1.60	0.45	-	5.60	0.41	
22	Spike AF 5%	AlloU, 24h, F	13.07	3.66	31; 18	2.00	73.60	5.34
23	NC	UCB, 0h	1.10	0.31	-	2.0	0.15	
24	Spike AF 5%	UCB, 0h	3.04	0.85	3; 1	0.65	9.7	0.70
25	NC	UCB, 4h, RT	1.63	0.46	-	4.25	0.31	
26	Spike AF 5%	UCB, 4h, RT	1.72	0.48	0	0	3.15	0.23

	Ref.						
27	Standard 0.1%	***	2.8	1; 0	0	25.35	1.84
28	Standard 1%	***	2.8	3; 5	1.0	22.95	1.67
29	Standard 5%	***	2.8	34; 23	5.1	25.90	1.88

NA*: sample could not be analyzed due to insufficient droplet generation.

**Calculated as mentioned above: $\frac{ng}{assay} = \frac{copies}{\mu l} * 22 \mu l * 0.0033 ng$.

*** Reference standard solutions were measured fluorometrically using Qbit resulting in 400ng/ml each.